



**LABORATORY
ACCREDITATION
BUREAU** a division of A-S-B

Certificate of Accreditation

ISO/IEC 17025:2005

Certificate Number L1147.10-1

FARO Business Technologies India Pvt. Ltd.
B-1, D-5 Mohan Cooperative Industrial Estate, Mathura Road
New Delhi – 110044 India

has met the requirements set forth in L-A-B's policies and procedures, all requirements of ISO/IEC 17025:2005 "General Requirements for the competence of Testing and Calibration Laboratories".*

The accredited lab has demonstrated technical competence to a defined "Scope of Accreditation" and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).

Accreditation valid through: January 25, 2017

R. Douglas Leonard, Jr., President, COO
Laboratory Accreditation Bureau
Presented the 13th of January 2014

*See the laboratory's Scope of Accreditation for details of accredited parameters

**Laboratory Accreditation Bureau is found to be in compliance with ISO/IEC 17011:2004 and recognized by ILAC (International Laboratory Accreditation Cooperation) and NACLA (National Cooperation for Laboratory Accreditation).

Scope of Accreditation For FARO Business Technologies India Pvt. Ltd.

B-1, D-5 Mohan Cooperative Industrial Estate, Mathura Road
New Delhi – 110044 India
Vishal Verma
91-11-4167-6330

In recognition of a successful assessment to ISO/IEC 17025:2005, accreditation is granted to **FARO Business Technologies India Pvt. Ltd.** to perform the following Calibrations:

Accreditation granted through: **January 25, 2017**

Calibration

Length – Hand Tools and Precision Gages 3D

Calibration Parameter/Equipment ¹	Range	Calibration and Measurement Capability(+/-)	Remarks
Articulated Arm Coordinate Measurement Machine Volumetric Performance ² (Ball Bar)	(0 to 2.2) m	(0.35 + 0.45L) μm	Articulated Arm Coordinate Measuring Machines (AACMM) produced by FARO Technologies, Inc.
Volumetric Performance ² (KSB Bar)	(0 to 2.2) m	3.5 μm	
Effective Diameter	(3 to 25.4) mm	1 μm	
Single Point Articulation Performance	N/A ³	0.41 μm	

Calibration and Measurement Capability represents expanded uncertainties at approximately a 95% confidence level using a coverage factor of k=2.

Notes:

- 1) Laboratory offers calibration services at the laboratory's own facilities.
- 2) *L* = length in meters
- 3) Point measurements do not have a range.

Approved by: 
R. Douglas Leonard
Chief Technical Officer

Date: January 13, 2014